



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 13, 2012 INITIALS: gjk OWNER: THURSTON COUNTY SHERIFF

DISTRICT: ADDRESS:

LOCATION: FM SHOP

COUNTY:

MAKE: HAENNI

TYPE: PORT. WHEEL SCALE

ACCURACY CLASS: IIII

MODEL: 101

ID. NO: 23168

CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs.

REG. ELEMENT: DIAL

PLATFORM 15"X26"


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19900	-100				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration",
Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan LaFreniere Date: May 14, 2012

Print name: DAN LAFRENIERE



State of Washington

County of Thurston

Signed and sworn to (or affirmed before
me on May 14, 2012 by

Patricia A. Geppert
(Signature of notary public)

Patricia A. Geppert
Printed Name of Notary

My appointment expires Sept. 4, 2015